



ZAVOD ZA HITNU MEDICINU KOPRIVNIČKO-KRIŽEVAČKE ŽUPANIJE

Na temelju članka 32. Statuta Zavoda za hitnu medicinu Koprivničko-križevačke županije, (KLASA: 003-01/16-01/01, URBROJ: 2137-89-16-01) od 27. srpnja 2016. i Suglasnosti o pokretanju prodaje putem javne prodaje dva hematološka analizatora Zavoda za hitnu medicinu Koprivničko-križevačke županije (KLASA: 024-01/17-01/02, URBROJ: 2137-89-17-11) od 28. veljače 2017. godine, ravnateljica Mirjana Hanžeković, mag.oec. objavljuje

NATJEČAJ **za prodaju dva hematološka analizatora** **Zavoda za hitnu medicinu** **Koprivničko-križevačke županije**

I.

Predmet natječaja je prodaja dva hematološka analizatora Zavoda za hitnu medicinu Koprivničko-križevačke županije.

II.

Procijenjena vrijednost:

Procijenjena vrijednost jednog hematološkog analizatora (s PDV-om) iznosi 26.778,75 kuna.

III.

Hematološki analizator
Specifikacije:

Sample volume	25 μ l of whole blood in normal 3-part or non-diff mode 50 μ l of whole blood in pre-diluted mode
Chambers	1 unified chamber for diluting whole blood and counting
Reagent System	Isotonic Diluent, Lyse, Cleaner
Aperture diameter	70 μ m (RBC/PLT, WBC)
Throughput	30 tests/hour

Characteristics Parameter	Accuracy	Reproducibility (CV)	Carry-over sample to sample	Test range	Unit
WBC	3%	3%	<1%	4.0-20.0	10 ³ / μ l
RBC	3%	2%	<1%	4.0-15.0	10 ⁶ / μ l
HCT	3%	3%	<1%	25.0-50.0	%
MCV	2%	1%	N/A	60-100	fl
HGB	2%	2%	<1%	9-16	g/dl
PLT	5%	5%	<3% or <20	200-900	10 ³ / μ l

Trg dr. Tomislava Bardeka 10, 48000 Koprivnica, tel/fax: 048/641 203, 641281, www.hitna-kckz.hr

IBAN:HR 4623860021110006914, Podravska banka Koprivnica OIB: 63076865469

Ustanova upisana kod Trgovačkog suda u Varaždinu. Osnivač Koprivničko-križevačka županija

Ravnateljica – Mirjana Hanžeković, mag.oec.



ZAVOD ZA HITNU MEDICINU KOPRIVNIČKO-KRIŽEVAČKE ŽUPANIJE

S	Sample volume	25 µl of whole blood in normal 3-part or non-diff mode 50 µl of whole blood in pre-diluted mode
S	Chambers	1 unified chamber for diluting whole blood and counting
S	Reagent system	Isotonic Diluent, Lyse, Cleaner
C	Aperture diameter	70 µm (RBC/PLT, WBC)
C	Throughput	30 tests/hour

Characteristics Parameter:	Accuracy	Reproducibility (CV)	Carry-over sample to sample	Test range
WBC	3%	3%	<1%	4.0-20.0
RBC	3%	2%	<1%	4.0-15.0
HCT	3%	3%	<1%	25.0-50.0
MCV	2%	1%	N/A	60-100
HGB	2%	2%	<1%	9-16
PLT	5%	5%	<3% or <20	200-900

Note: Differential results for WBC counts below 1.00 are not displayed, d

I	Sampling method	Open tube system with automatic sample rotor.
I	Sample types	Human (general), male, female, baby, toddler and child.
E	Clog prevention	High-voltage pulse on aperture in each analysis cycle, chemical high pressure back-flush of the aperture using Cleaner reagent.
E	Cleaning procedure	High-voltage burst of the aperture, high-pressure back-flush, chemical cleaning of the aperture using Cleaner reagent.
I	Calibration	one or more measurement automatic and manual (factors) calibration WBC, HGB, RBC, PLT, MCV (or HCT), RDW, and MPV.
S	User interface	Easy-to-use, menu driven user interface with touch-screen and START button, status LED.
P	Languages available	Korean, Chinese, Spanish, Vietnamese, Polish, German, French, Portuguese, Russian, Italian, Czech, Greek and English
B	Data capacity	10 000 results, with RBC, PLT, and WBC 3-part histogram
L	Host computer interface	USB B port
L	Data back-up method	USB mass storage device (PenDrive™)
U	Software upgrade method	via USB A port using USB mass storage device (PenDrive™)
U	Printer interface	USB with support for HP printers (DeskJet, LaserJet, PCL3, PCL6), EPSON Dot-matrix printers
E	Built-in printer	Axiom thermal printer module, 58 mm wide roll paper, full report histograms
P	Display	320x240 -dots, high-contrast, backlit, color graphics LCD (liquid crystal display)
P	User interface	Full-LCD Touch-screen + separate START button, red/green status LEDs
C	External keyboard	USB keyboard via USB A port
C	Power requirement	12VDC, 5A, 60W max. operating power
S	Power supply unit	External, auto-ranging power unit for 100-120 or 200-240 VAC
S	Operating temperature	59–86 °F (15–30 °C). Optimal temperature is 77 °F (25 °C)
S	Storage temperature	41-95 °F (5-35 °C)
S	Transport temperature	41-104 °F (5-40 °C)
S	Dimensions (W x D x H)	12.6 x 10.2 x 14.4 in (320 x 260 x 365 mm)
S	Net weight	12 kg

Sample volume 25 µl of whole blood in normal 3-part or non-diff mode



ZAVOD ZA HITNU MEDICINU KOPRIVNIČKO-KRIŽEVAČKE ŽUPANIJE

Transport temperature	41-104 °F (5-40 °C)
Dimensions	12.6 x 10.2 x 14.4 in (320 x 260 x 365 mm)
Net weight	12 kg

IV.

Pravo sudjelovanja u natječaju imaju sve fizičke osobe sa prebivalištem u RH i pravne osobe sa sjedištem u RH.

Ponuda mora sadržavati: ime i prezime, odnosno naziv pravne osobe, adresu, OIB i ponuđenu cijenu.

Ponuda se dostavlja u pisanom obliku, zatvorenoj omotnici s naznakom „NATJEČAJ-PONUDA ZA KUPNJU HEMATOLOŠKOG ANALIZATORA“ na adresu prodavatelja: Zavod za hitnu medicinu Koprivničko-križevačke županije, Trg dr. Tomislava Bardeka 10, 48000 Koprivnica u roku od 15 dana od dana objave natječaja.

V.

Ponuda pristigla poslije roka za dostavu ponuda neće se otvoriti nego će se odmah vratiti ponuditelju.

Svi ponuditelji biti će obaviješteni o rezultatima natječaja najkasnije u roku od 8 dana od dana isteka roka za dostavu ponuda.

Ako najpovoljniji ponuditelj odustane od kupnje hematološkog analizatora, prodavatelj će izabrati sljedećeg najpovoljnijeg ponuditelja.

Osoba za kontakt: Mirjana Hanžeković, 048/641-250.

VI.

Najpovoljnijom ponudom smatrati će se ponuda s najvišom ponuđenom cijenom.

U slučaju dviju ili više ponuda s istom cijenom, odabrat će se ranije zaprimljena ponuda.

Kupac može preuzeti hematološki analizator hitne medicinske pomoći čim uplata bude vidljiva na žiro-računu Zavoda za hitnu medicinu Koprivničko-križevačke županije.

KLASA: 406-01/17-01/12

URBROJ: 2137-89-17-01

U Koprivnici, 11. travnja 2017. godine

RAVNATELJICA:
Mirjana Hanžeković, mag.oec.

Trg dr. Tomislava Bardeka 10, 48000 Koprivnica, tel/fax: 048/641 203, 641281, www.hitna-kckz.hr

IBAN:HR 4623860021110006914, Podravska banka Koprivnica OIB: 63076865469

Ustanova upisana kod Trgovačkog suda u Varaždinu. Osnivač Koprivničko-križevačka županija

Ravnateljica – Mirjana Hanžeković, mag.oec.